

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (M.P.E)</b>	<b>Incertidumbre (Uncertainty)</b>	<b>Regla de decisión</b>
100.0004 ml	100 ml	-0.0004 ml	± 1.0 ml	± 3.1E+00 µl	Cumple
150.0448 ml	150 ml	-0.0448 ml	± 1.0 ml	± 4.6E+00 µl	Cumple
200.0231 ml	200 ml	-0.0231 ml	± 1.0 ml	± 6.1E+00 µl	Cumple