

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	Regla de decisión
99.9885 ml	100 ml	0.0115 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
249.9607 ml	250 ml	0.0393 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
499.9694 ml	500 ml	0.0306 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
749.9261 ml	750 ml	0.0739 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
999.9108 ml	1000 ml	0.0892 ml	± 5.0 ml	± 3.5E-01 µl	Cumple