

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (M.P.E) | Incertidumbre (Uncertainty) | Regla de decisión |
|--------------------------|------------------|----------------------|----------------------|------------------------------------|--------------------------|
| 99.9955 ml | 100 ml | 0.0045 ml | ± 5.0 ml | ± 3.5E-01 µl | Cumple |
| 249.9316 ml | 250 ml | 0.0684 ml | ± 5.0 ml | ± 3.5E-01 µl | Cumple |
| 499.9754 ml | 500 ml | 0.0246 ml | ± 5.0 ml | ± 3.5E-01 µl | Cumple |
| 749.9160 ml | 750 ml | 0.0840 ml | ± 5.0 ml | ± 3.5E-01 µl | Cumple |
| 999.9121 ml | 1000 ml | 0.0879 ml | ± 5.0 ml | ± 3.5E-01 µl | Cumple |