

| <b>Patrón (Standard)</b> | <b>UBP (UUT)</b> | <b>Error (Error)</b> | <b>E.M.P (M.P.E)</b> | <b>Incertidumbre (Uncertainty)</b> | <b>Regla de decisión</b> |
|--------------------------|------------------|----------------------|----------------------|------------------------------------|--------------------------|
| 100.0085 ml              | 100 ml           | -0.0085 ml           | ± 5.0 ml             | ± 3.5E-01 µl                       | Cumple                   |
| 249.9406 ml              | 250 ml           | 0.0594 ml            | ± 5.0 ml             | ± 3.5E-01 µl                       | Cumple                   |
| 499.9965 ml              | 500 ml           | 0.0035 ml            | ± 5.0 ml             | ± 3.5E-01 µl                       | Cumple                   |
| 749.9335 ml              | 750 ml           | 0.0665 ml            | ± 5.0 ml             | ± 3.5E-01 µl                       | Cumple                   |
| 999.9018 ml              | 1000 ml          | 0.0982 ml            | ± 5.0 ml             | ± 3.5E-01 µl                       | Cumple                   |