

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	Regla de decisión
100.0035 ml	100 ml	-0.0035 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
249.9857 ml	250 ml	0.0143 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
499.9944 ml	500 ml	0.0056 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
750.0599 ml	750 ml	-0.0599 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
999.9100 ml	1000 ml	0.0900 ml	± 5.0 ml	± 3.5E-01 µl	Cumple