

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (M.P.E)</b>	<b>Incertidumbre (Uncertainty)</b>	<b>Regla de decisión</b>
99.9935 ml	100 ml	0.0065 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
249.9857 ml	250 ml	0.0143 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
499.9804 ml	500 ml	0.0196 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
749.9300 ml	750 ml	0.0700 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
999.9172 ml	1000 ml	0.0828 ml	± 5.0 ml	± 3.5E-01 µl	Cumple