

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	Regla de decisión
99.9936 ml	100 ml	0.0064 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
249.9977 ml	250 ml	0.0023 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
500.0198 ml	500 ml	-0.0198 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
749.9150 ml	750 ml	0.0850 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
999.9100 ml	1000 ml	0.0900 ml	± 5.0 ml	± 3.5E-01 µl	Cumple