

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	Regla de decisión
100.0075 ml	100 ml	-0.0075 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
249.9927 ml	250 ml	0.0073 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
499.9955 ml	500 ml	0.0045 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
750.0602 ml	750 ml	-0.0602 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
999.9288 ml	1000 ml	0.0712 ml	± 5.0 ml	± 3.5E-01 µl	Cumple