

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	Regla de decisión
99.9913 ml	100 ml	0.0087 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
249.9085 ml	250 ml	0.0915 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
499.9073 ml	500 ml	0.0927 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
750.0213 ml	750 ml	-0.0213 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
999.9950 ml	1000 ml	0.0050 ml	± 5.0 ml	± 3.5E-01 µl	Cumple