

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (M.P.E)</b>	<b>Incertidumbre (Uncertainty)</b>	<b>Regla de decisión</b>
99.9865 ml	100 ml	0.0135 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
249.9005 ml	250 ml	0.0995 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
499.9213 ml	500 ml	0.0787 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
750.0012 ml	750 ml	-0.0012 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
999.9288 ml	1000 ml	0.0712 ml	± 5.0 ml	± 3.5E-01 µl	Cumple