

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (M.P.E)</b>	<b>Incertidumbre (Uncertainty)</b>	<b>Regla de decisión</b>
100.0138 ml	100 ml	-0.0138 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
250.0043 ml	250 ml	-0.0043 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
500.0079 ml	500 ml	-0.0079 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
749.9114 ml	750 ml	0.0886 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
999.9055 ml	1000 ml	0.0945 ml	± 5.0 ml	± 3.5E-01 µl	Cumple