

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (M.P.E)</b>	<b>Incertidumbre (Uncertainty)</b>	<b>Regla de decisión</b>
100.0286 ml	100 ml	-0.0286 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
249.9516 ml	250 ml	0.0484 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
499.9855 ml	500 ml	0.0145 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
749.9912 ml	750 ml	0.0088 ml	± 5.0 ml	± 3.5E-01 µl	Cumple
999.9218 ml	1000 ml	0.0782 ml	± 5.0 ml	± 3.5E-01 µl	Cumple