

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
50.10 mmHg	50 mmHg	-0.10 mmHg	± 3 mmHg	±	6.6E-01 mmHg
100.10 mmHg	100 mmHg	-0.10 mmHg	± 3 mmHg	±	6.5E-01 mmHg
149.90 mmHg	150 mmHg	0.10 mmHg	± 3 mmHg	±	6.6E-01 mmHg
200.00 mmHg	200 mmHg	0.00 mmHg	± 3 mmHg	±	6.8E-01 mmHg
249.90 mmHg	250 mmHg	0.10 mmHg	± 3 mmHg	±	7.3E-01 mmHg
300.10 mmHg	300 mmHg	-0.10 mmHg	± 3 mmHg	±	7.3E-01 mmHg