

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
50.00 mmHg	50 mmHg	0.00 mmHg	± 3 mmHg	±	6.6E-01 mmHg
99.90 mmHg	100 mmHg	0.10 mmHg	± 3 mmHg	±	6.5E-01 mmHg
149.90 mmHg	150 mmHg	0.10 mmHg	± 3 mmHg	±	6.6E-01 mmHg
200.00 mmHg	200 mmHg	0.00 mmHg	± 3 mmHg	±	6.8E-01 mmHg
250.00 mmHg	250 mmHg	0.00 mmHg	± 3 mmHg	±	7.3E-01 mmHg
299.90 mmHg	300 mmHg	0.10 mmHg	± 3 mmHg	±	7.3E-01 mmHg