

| Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) | |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|--------------|
| 50.00 mmHg | 50 mmHg | 0.00 mmHg | ± 3 mmHg | ± | 6.6E-01 mmHg |
| 100.00 mmHg | 100 mmHg | 0.00 mmHg | ± 3 mmHg | ± | 6.5E-01 mmHg |
| 150.10 mmHg | 150 mmHg | -0.10 mmHg | ± 3 mmHg | ± | 6.6E-01 mmHg |
| 200.10 mmHg | 200 mmHg | -0.10 mmHg | ± 3 mmHg | ± | 6.8E-01 mmHg |
| 249.90 mmHg | 250 mmHg | 0.10 mmHg | ± 3 mmHg | ± | 7.3E-01 mmHg |
| 300.00 mmHg | 300 mmHg | 0.00 mmHg | ± 3 mmHg | ± | 7.3E-01 mmHg |