

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.423 °C	30.6 °C	0.177 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.609 %HR	32.1 %HR	2.491 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.640 %HR	51.9 %HR	2.260 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.090 %HR	75.1 %HR	0.010 %HR	± 3.0 %HR	± 1.9E+00 %HR