

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.423 °C	30.4 °C	-0.023 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.609 %HR	32.5 %HR	2.891 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.640 %HR	50.1 %HR	0.460 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.090 %HR	73.5 %HR	-1.590 %HR	± 3.0 %HR	± 1.9E+00 %HR