

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.5	0.49909	0.910	± 1.6	± 4.3E-01 mm					
0 a 1	0.99906	0.940	± 2	± 4.3E-01 mm					
0 a 1.5	1.49911	0.890	± 2.4	± 4.8E-01 mm					
0 a 2	1.99936	0.640	± 2.8	± 4.8E-01 mm					
0 a 2.5	2.49898	1.020	± 3.2	± 4.8E-01 mm					
0 a 3	2.99900	1.000	± 3.6	± 4.8E-01 mm					
0 a 3.5	3.49922	0.780	± 4	± 4.8E-01 mm					
0 a 4	3.99908	0.920	± 4.4	± 4.8E-01 mm					
0 a 4.5	4.49978	0.220	± 4.8	± 4.8E-01 mm					
0 a 5	4.99935	0.650	± 5.2	± 4.8E-01 mm					
