

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.175 °C	20.8 °C	0.625 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.109 °C	25.6 °C	0.491 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.012 °C	30.6 °C	0.588 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.519 %HR	44.8 %HR	14.281 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.537 %HR	60.2 %HR	10.663 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.783 %HR	79.8 %HR	10.017 %HR	± 3.0 %HR	± 1.9E+00 %HR