

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	20.175 °C	19.9 °C	-0.275 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.109 °C	24.8 °C	-0.309 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.012 °C	29.5 °C	-0.512 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.519 %HR	46.6 %HR	16.081 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.537 %HR	62.7 %HR	13.163 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.783 %HR	82.0 %HR	12.217 %HR	± 3.0 %HR	± 1.9E+00 %HR