

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.012 °C	30.4 °C	0.388 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.519 %HR	30.0 %HR	-0.519 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.537 %HR	47.0 %HR	-2.537 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.895 %HR	77.0 %HR	2.105 %HR	± 5.0 %HR	± 1.9E+00 %HR