

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
30 °C	30.012 °C	29.9 °C	-0.112 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.519 %HR	32.9 %HR	2.381 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.537 %HR	51.2 %HR	1.663 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.895 %HR	73.6 %HR	-1.295 %HR	± 3.0 %HR	± 1.9E+00 %HR