

| <b>Nominal</b> ( <i>Nominal</i> ) | <b>Patrón</b> ( <i>Standard</i> ) | <b>UBP</b> ( <i>UUT</i> ) | <b>Error</b> ( <i>Error</i> ) | <b>Exactitud</b> ( <i>Accuracy</i> ) | <b>Incertidumbre</b> ( <i>Uncertainty</i> ) |
|-----------------------------------|-----------------------------------|---------------------------|-------------------------------|--------------------------------------|---|
| 30 °C                             | 30.012 °C                         | 29.9 °C                   | -0.112 °C                     | ± 0.5 °C                             | ± 4.0E-01 °C                                |
| 30 %HR                            | 30.519 %HR                        | 32.7 %HR                  | 2.181 %HR                     | ± 3.0 %HR                            | ± 1.9E+00 %HR                               |
| 50 %HR                            | 49.537 %HR                        | 51.2 %HR                  | 1.663 %HR                     | ± 3.0 %HR                            | ± 1.9E+00 %HR                               |
| 75 %HR                            | 74.895 %HR                        | 73.9 %HR                  | -0.995 %HR                    | ± 3.0 %HR                            | ± 1.9E+00 %HR                               |