

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.905 °C	2.1 °C	0.195 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	3.900 °C	4.1 °C	0.200 °C	± 0.5 °C	± 4.0E-01 °C
6 °C	5.856 °C	6.0 °C	0.144 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.891 °C	8.1 °C	0.209 °C	± 0.5 °C	± 4.0E-01 °C