

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.905 °C	2.1 °C	0.195 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	3.900 °C	4.1 °C	0.200 °C	± 0.5 °C	± 4.0E-01 °C
6 °C	5.856 °C	6.0 °C	0.144 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.891 °C	8.0 °C	0.109 °C	± 0.5 °C	± 4.0E-01 °C