

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.012 °C	29.8 °C	-0.212 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.519 %HR	32.6 %HR	2.081 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.537 %HR	51.9 %HR	2.363 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.895 %HR	74.4 %HR	-0.495 %HR	± 3.0 %HR	± 1.9E+00 %HR