

Certificado No. (Certificate #):

68141

Equipo (Instrument): Calibrador Pie de Rey

Fecha de Calibración (Calibration Date):

2026-04-15

Marca (Brand):

Mitutoyo

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	9.40 mm	9.40 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Exterior / (Exterior)	38.10 mm	38.10 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Exterior / (Exterior)	100.00 mm	100.00 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Exterior / (Exterior)	125.40 mm	125.41 mm	10 μm	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Exterior / (Exterior)	147.50 mm	147.52 mm	20 μm	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Interior / (Interior)	76.20 mm	76.20 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 8.2\text{E}+00 \mu\text{m}$
Interior / (Interior)	100.00 mm	100.00 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 8.2\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	28.70 mm	28.70 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 7.1\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	100.00 mm	100.00 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 7.1\text{E}+00 \mu\text{m}$