

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (Tolerance)</b>	<b>Incertidumbre (Uncertainty)</b>	
52.65 $\mu\text{m}$	52.70 $\mu\text{m}$	0.05 $\mu\text{m}$	$\pm 1.6 \mu\text{m}$	$\pm$	5.8E-02 $\mu\text{m}$
155.23 $\mu\text{m}$	155.7 $\mu\text{m}$	0.47 $\mu\text{m}$	$\pm 4.7 \mu\text{m}$	$\pm$	6.2E-02 $\mu\text{m}$
257.12 $\mu\text{m}$	257.1 $\mu\text{m}$	-0.02 $\mu\text{m}$	$\pm 7.7 \mu\text{m}$	$\pm$	7.0E-02 $\mu\text{m}$
511.72 $\mu\text{m}$	511.9 $\mu\text{m}$	0.18 $\mu\text{m}$	$\pm 15.4 \mu\text{m}$	$\pm$	9.7E-02 $\mu\text{m}$