

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	29.782 °C	29.3 °C	-0.482 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.762 %HR	31 %HR	1.238 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.783 %HR	49 %HR	-0.783 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.028 %HR	67 %HR	-3.028 %HR	± 5.0 %HR	± 1.9E+00 %HR