

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.782 °C	29.3 °C	-0.482 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.762 %HR	33 %HR	3.238 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.783 %HR	50 %HR	0.217 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.028 %HR	68 %HR	-2.028 %HR	± 5.0 %HR	± 1.9E+00 %HR