

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.782 °C	30.2 °C	0.418 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.762 %HR	34 %HR	4.238 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.783 %HR	52 %HR	2.217 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.028 %HR	71 %HR	0.972 %HR	± 5.0 %HR	± 1.9E+00 %HR