

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.782 °C	29.5 °C	-0.282 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.762 %HR	32.6 %HR	2.838 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.783 %HR	52.2 %HR	2.417 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.028 %HR	68.1 %HR	-1.928 %HR	± 3.0 %HR	± 1.9E+00 %HR