

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.095 °C	2.9 °C	0.805 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	8.062 °C	8.8 °C	0.738 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	29.902 °C	30.2 °C	0.298 °C	± 1.0 °C	± 4.0E-01 °C
70 %HR	69.537 %HR	69.8 %HR	0.263 %HR	± 5.0 %HR	± 1.9E+00 %HR