

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
2 °C	2.095 °C	3.0 °C	0.905 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	8.062 °C	8.9 °C	0.838 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	29.902 °C	30.5 °C	0.598 °C	± 1.0 °C	± 4.0E-01 °C
70 %HR	69.537 %HR	69.7 %HR	0.163 %HR	± 5.0 %HR	± 1.9E+00 %HR