

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.095 °C	1.9 °C	-0.195 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.062 °C	7.9 °C	-0.162 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.902 °C	29.9 °C	-0.002 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	69.537 %HR	70.2 %HR	0.663 %HR	± 3.0 %HR	± 1.9E+00 %HR