

Patrón (Standard)	UBP (UT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
3600.009 s	3600 s	-0.009 s	-0.0002%	± 0.01 %	± 1.0E+03 s
3600.014 s	3600 s	-0.014 s	-0.0004%	± 0.01 %	± 1.0E+03 s
3600.015 s	3600 s	-0.015 s	-0.0004%	± 0.01 %	± 1.0E+03 s