



Certificado No.:

63333

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-11-15

\*\*\*\*\*

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
8:30:00 AM	15.8	16.0	0.20
8:31:00 AM	15.8	16.4	0.60
8:32:00 AM	15.7	15.8	0.10
8:33:00 AM	15.7	15.7	0.00
8:34:00 AM	15.6	15.6	0.00
8:35:00 AM	15.4	15.4	0.00
8:36:00 AM	15.3	15.0	-0.30
8:37:00 AM	15.3	15.0	-0.30
8:38:00 AM	15.5	14.9	-0.60
8:39:00 AM	15.7	14.8	-0.90
8:40:00 AM	16.0	14.7	-1.30
8:41:00 AM	16.0	14.6	-1.40
8:42:00 AM	16.1	14.5	-1.60
8:43:00 AM	16.3	14.4	-1.90
8:44:00 AM	16.4	14.3	-2.10
8:45:00 AM	16.5	14.0	-2.50
8:46:00 AM	16.6	14.0	-2.60
8:47:00 AM	16.7	14.2	-2.50
8:48:00 AM	16.4	14.3	-2.10
8:49:00 AM	16.3	14.3	-2.00
8:50:00 AM	16.2	14.5	-1.70
8:51:00 AM	16.1	15.6	-0.50
8:52:00 AM	15.9	14.4	-1.50
8:53:00 AM	15.8	14.1	-1.70
8:54:00 AM	15.7	14.0	-1.70
8:55:00 AM	15.4	14.0	-1.40
8:56:00 AM	15.3	13.9	-1.40
8:57:00 AM	15.2	13.8	-1.40
8:58:00 AM	15.1	13.7	-1.40
8:59:00 AM	15.0	13.6	-1.40
<b>Promedio</b>	<b>15.83</b>	<b>14.65</b>	
<b>Desviación</b>	<b>0.466</b>	<b>0.744</b>	
<b>Max</b>	<b>16.70</b>	<b>16.40</b>	
<b>Min</b>	<b>15.00</b>	<b>13.70</b>	

Certificado No.:

63333

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-11-15

\*\*\*\*\*

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
8:30:00 AM	60.4	60.8	0.40
8:31:00 AM	60.6	59.8	-0.80
8:32:00 AM	60.6	59.7	-0.90
8:33:00 AM	60.7	59.5	-1.20
8:34:00 AM	60.8	59.4	-1.40
8:35:00 AM	60.9	59.3	-1.60
8:36:00 AM	61.0	59.2	-1.80
8:37:00 AM	61.8	59.0	-2.80
8:38:00 AM	61.5	59.0	-2.50
8:39:00 AM	61.7	58.8	-2.90
8:40:00 AM	61.6	58.7	-2.90
8:41:00 AM	61.4	58.7	-2.70
8:42:00 AM	61.7	58.6	-3.10
8:43:00 AM	61.9	58.5	-3.40
8:44:00 AM	61.7	58.4	-3.30
8:45:00 AM	62.4	58.3	-4.10
8:46:00 AM	62.5	58.9	-3.60
8:47:00 AM	62.2	59.2	-3.00
8:48:00 AM	62.0	59.3	-2.70
8:49:00 AM	61.8	59.4	-2.40
8:50:00 AM	61.7	59.5	-2.20
8:51:00 AM	61.7	59.7	-2.00
8:52:00 AM	61.5	59.8	-1.70
8:53:00 AM	61.4	60.4	-1.00
8:54:00 AM	61.0	60.5	-0.50
8:55:00 AM	60.9	60.7	-0.20
8:56:00 AM	60.8	60.8	0.00
8:57:00 AM	60.5	61.2	0.70
8:58:00 AM	60.4	61.4	1.00
8:59:00 AM	60.3	61.7	1.40
<b>Promedio</b>	<b>61.31</b>	<b>59.61</b>	
<b>Desviación</b>	<b>0.627</b>	<b>0.933</b>	
<b>Max</b>	<b>62.50</b>	<b>61.40</b>	
<b>Min</b>	<b>60.30</b>	<b>58.30</b>	

Certificado No.:

63333

Registrador de Temperatura y Humedad


Fecha de Calibración:

2025-11-15

\*\*\*\*\*

### Curva de Calibración - TEMPERATURA

error ± Incertidumbre



### Curva de Calibración - % HUMEDAD RELATIVA

error ± Incertidumbre

