

Certificado No.: 60386
 Fecha de Calibración: 2025-08-22

Equipo: Registrador de Temperatura
 Marca: QUECKLINC

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
12:32:00 PM	13.30	13.3	0.00
12:33:00 PM	13.10	13.0	-0.10
12:34:00 PM	12.90	12.7	-0.20
12:35:00 PM	12.70	12.5	-0.20
12:36:00 PM	12.30	12.1	-0.20
12:37:00 PM	12.10	11.8	-0.30
12:38:00 PM	11.80	11.5	-0.30
12:39:00 PM	11.60	11.3	-0.30
12:40:00 PM	11.30	11.0	-0.30
12:41:00 PM	11.00	10.8	-0.20
12:42:00 PM	10.70	10.5	-0.20
12:43:00 PM	10.40	10.2	-0.20
12:44:00 PM	10.20	10.0	-0.20
12:45:00 PM	10.00	9.8	-0.20
12:46:00 PM	9.80	9.6	-0.20
12:47:00 PM	9.60	9.4	-0.20
12:48:00 PM	9.40	9.2	-0.20
12:49:00 PM	9.20	9.1	-0.10
12:50:00 PM	9.00	8.8	-0.20
12:51:00 PM	8.80	8.6	-0.20
12:52:00 PM	8.60	8.4	-0.20
12:53:00 PM	8.40	8.3	-0.10
12:54:00 PM	8.20	8.0	-0.20
12:55:00 PM	8.00	7.8	-0.20
12:56:00 PM	7.80	7.7	-0.10
12:57:00 PM	7.60	7.5	-0.10
12:58:00 PM	7.40	7.3	-0.10
12:59:00 PM	7.20	7.1	-0.10
1:00:00 PM	7.10	6.9	-0.20
1:01:00 PM	7.00	6.9	-0.10
Promedio	9.88	9.7	
Desviación	1.971	1.952	
Max	13.30	13.3	
Min	7.00	6.9	

Certificado No.: 60386
Fecha de Calibración: 2025-08-22

Equipo: Registrador de Temperatura
Marca: QUECKLINC

