

Certificado No.:

61416

Controlador de Temperatura + Sensor

Fecha de Calibración:

2025-09-17

ELIWELL

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (175 °C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
1:20:00 PM	-22.81	-25	-2.19
1:21:00 PM	-23.21	-25	-1.79
1:22:00 PM	-23.48	-25	-1.52
1:23:00 PM	-23.55	-25	-1.45
1:24:00 PM	-24.06	-26	-1.94
1:25:00 PM	-24.34	-26	-1.66
1:26:00 PM	-24.51	-26	-1.49
1:27:00 PM	-24.72	-26	-1.28
1:28:00 PM	-24.89	-27	-2.11
1:29:00 PM	-25.10	-27	-1.90
1:30:00 PM	-25.28	-27	-1.72
1:31:00 PM	-25.39	-27	-1.61
1:32:00 PM	-25.67	-27	-1.33
1:33:00 PM	-25.81	-28	-2.19
1:34:00 PM	-26.00	-28	-2.00
1:35:00 PM	-26.13	-28	-1.87
1:36:00 PM	-26.27	-28	-1.73
1:37:00 PM	-26.40	-28	-1.60
1:38:00 PM	-26.54	-28	-1.46
1:39:00 PM	-26.72	-28	-1.28
1:40:00 PM	-26.90	-29	-2.10
1:41:00 PM	-26.98	-29	-2.02
1:42:00 PM	-27.15	-29	-1.85
1:43:00 PM	-27.26	-29	-1.74
1:44:00 PM	-27.45	-29	-1.55
1:45:00 PM	-27.56	-29	-1.44
1:46:00 PM	-27.68	-29	-1.32
1:47:00 PM	-27.80	-29	-1.20
1:48:00 PM	-27.90	-29	-1.10
1:49:00 PM	-28.41	-30	-1.59
Promedio	-25.87	-27.5	
Desviación	1.542	1.479	
Max	-22.81	-25	
Min	-28.41	-29	

Certificado No.:

61416

Controlador de Temperatura + Sensor

Fecha de Calibración:

2025-09-17

ELIWELL

